

MLS Cold Laser Settings

Handheld Applicator Settings				
Pathology	Frequency (Hz)	Time	Intensity (%)	Energy (J)
Sinusitis	100	2:00	50	31.38
Trigeminal Neuralgia	100	2:00	50	31.38
Vasomotor Rhinitis	100	2:00	50	31.38
Classical Headache	400	2:00	50	35.52
Migraine	400	2:00	50	35.52
Acne	500	2:00	50	36.91
Brachial Neuralgia	500	2:00	50	36.91
Dislocation	500	2:00	50	36.91
Hand/Foot Arthritis	500	2:00	50	36.91
Contusion	700	2:00	50	39.67
Knee Arthrosis	700	2:00	50	39.67
Lumbago	700	2:00	50	39.67
Muscular Contraction	700	2:00	50	39.67
Sprain	700	2:00	50	39.67
Tendinitis	700	2:00	50	39.67
Cervical Arthrosis	700	3:00	50	59.51
Epicondylitis	700	3:00	50	59.51
Muscle Sprain Inferior Limbs	700	3:00	50	59.51
Scapulo-Humeral Periarthritis	900	2:00	50	42.44
Backache	900	3:00	50	63.66
Lumbosciatica	900	3:00	50	63.66
Sciatic Neuralgia	900	3:00	50	63.66
Burn	1500	2:00	50	50.73
Dorsal Arthrosis	1500	2:00	50	50.73
Haematoma	1500	2:00	50	50.73
Muscular Tear	1500	2:00	50	50.73
Varix Ulcer	1500	2:00	50	50.73
Wound	1500	2:00	50	50.73
Hip Arthrosis	1500	3:00	50	76.10
Lumbosacral Athrosis	1500	3:00	50	76.10
Trigger Point Muscle	10	0:30	50	7.50
TMJ 30s open, 30s closed, 2x	350	1:00	50	17.33
Neck and TMJ muscle per side	700	4:00	50	79.34
Deprogramming (Dental)	900	3:00	50	63.66

Charlie Head Settings				
Frequency (Hz)	Time	Intensity (%)	Energy (J)	
100	2:00	50	90.9	
100	2:00	50	90.9	
100	2:00	50	90.9	
400	2:30	50	117	
400	2:00	50	93.59	
500	2:30	50	118.1	
500	4:00	50	189	
500	2:30	50	118.1	
500	2:30	50	118.1	
700	2:30	50	120.3	
700	3:30	50	168.5	
700	5:30	50	264.8	
700	3:30	50	168.5	
700	3:30	50	168.5	
700	2:30	50	120.3	
700	4:00	50	192.6	
700	3:00	50	144.4	
700	7:30	50	361.1	
900	4:00	50	196.1	
900	8:00	50	392.3	
900	10:00	50	490.3	
900	9:00	50	441.4	
1500	3:30	50	181.1	
1500	8:30	50	439.8	
1500	3:30	50	181.1	
1500	5:00	50	258.7	
1500	2:30	50	129.3	
1500	2:30	50	129.3	
1500	8:00	50	414	
1500	9:30	50	491.6	